

## ILLUMINA DECLARATION OF CONFORMITY

Illumina Inc. hereby declares under its sole responsibility that the product(s) listed are in conformity with the stipulations of the Essential Requirements of the EMC Directive [2004/108/EC] and the Low Voltage Directive [73/23/EEC amended by 93/68/EEC].

Manufacturer's Name: Illumina, Inc.  
25861 Industrial Blvd.  
Hayward, CA 94545  
USA  
Device Name: Genome Analyzer  
Device Commercial Name: Genome Analyzer Ii  
Genome Analyzer Iix  
Device Model: 1200 and 1206  
Manufacturing Site: Hayward CA  
CE Marking Granted: 2008

The reference product(s) and/or their components conform, at a minimum, to the following standard(s) and/or other normative document(s), pursuant to the provisions of the EMC Directive:

EN 61326-1:2006  
FCC Part 15, Subpart B  
Industry Canada ICES-003  
ANSI/UL 61010-1:2005, 2<sup>nd</sup> Edition  
CAN/CSA C22.2 61010-1:2004  
EN 61010-1:2001, 2<sup>nd</sup> Edition  
IEC 60825-1:2007, 2<sup>nd</sup> Edition

Authorized by:



Don Ellis

Sr. Director of Quality  
Illumina, Inc.

07/27/2010  
Date