



CERTIFICATE

No. U8 10 09 66290 006

Holder of Certificate: Illumina Inc.

illumina

9885 Towne Centre Drive
San Diego CA 92121
USA

Production Facility(ies): 66290

Certification Mark:



Product: Analysis Equipment (DNA Analyzer)

Model(s): SY-101-1001, SY-101-1002

Parameters:

Input Voltage:	100-120/200-240 VAC
Input Frequency:	50/60 Hz
Input Power:	360 VA
Protection Class:	I
Laser Class:	Class 1
Temperature, Ambient:	30°C maximum

Tested according to:

CAN/CSA-C22.2 No. 61010-1:2004
CAN/CSA-E60825-1:R2007
UL 61010-1/R:2008-10
EN 61010-1:2001, EN 60825-1:2007

The product was voluntarily tested according to the relevant safety requirements and mentioned properties. It can be marked with the certification mark shown above. The certification mark must not be altered in any way. This product certification system operated by TÜV SÜD America Inc. most closely resembles that described by ISO/IEC Guide 67, Conformity assessment - Fundamentals of product certification, System 3. See also notes overleaf.

Test report no.: 095-802279-200

Date: 2010-09-22

Page 1 of 1

