



## Declaration of Conformity

Illumina, Inc. hereby declares under its sole responsibility that the product(s) listed are in conformity to the EMC Directive [2004/108/EC], Low Voltage Directive [2006/95/EC], and R&TTE Directive [1999/5/EC].

MANUFACTURER: Illumina, Inc  
ADDRESS: 5200 Illumina Way  
San Diego, CA 92122, USA

FACTORY LOCATION: 25861 Industrial Blvd.  
Hayward, CA 94545, USA

PRODUCT TYPE: Genetic Sequencer  
MODEL: MiSeq, MiSeqDx, MiSeqFGx  
CE MARK AFFIXED: 2011

The construction of the product is in compliance with the following harmonized and/or consensus standards.

IEC/EN 61010-1:2001	<i>Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use - Part 1: General Requirements</i>
UL 61010-1:2004	
IEC/EN 61010-1:2010	
CAN/CSA 61010-1-4	
IEC 61010-2-081:2001+Amd1	<i>Particular requirements for automatic and semi-automatic laboratory equipment for analysis and other purposes</i>
IEC/EN 61010-2-101:2002	<i>Particular requirements for in vitro diagnostic (IVD) medical equipment</i>
EN 61326-1:2006 (Class A)	<i>Electrical equipment for the measurement, control and Laboratory use – EMC Requirements Part1, Class A</i>
EN 61326-2-6:2006	<i>EMC requirements -- Part 2-6: Particular requirements - In vitro diagnostic (IVD) medical equipment</i>
EN 50364:2001	<i>Limitation of human exposure to electromagnetic fields from devices operating in the frequency range 0 Hz to 300 GHz, used in electronic article surveillance (EAS), radio frequency identification (RFID) and similar applications</i>
ETSI EN 300 330-1 V1.7.1	<i>Electromagnetic compatibility and Radio spectrum Matters (ERM); Short Range Devices (SRD); Radio equipment in the frequency range 9 kHz to 25 MHz and inductive loop systems in the frequency range 9 kHz to 30 MHz;</i> <i>Part 1: Technical characteristics and test methods</i> <i>Part 2: Harmonized EN covering the essential requirements of article 3.2 of the R&amp;TTE Directive</i>
ETSI EN 300 330-2 V1.5.1	
Title 47, CFR Part 15 Subpart C § 15.225	<i>Intentional radiators operating within the band 13.110–14.010 MHz.</i>
Title 47, CFR Part 15 Subpart B	<i>Unintentional radio frequency devices</i>

Authorized by:

  
\_\_\_\_\_  
Mya Thomas  
VP, Regulatory Affairs

  
\_\_\_\_\_  
Date