

## Certificate of Analysis

### Description

Product Name	TruSeq Custom Amplicon Dx FFPE QC Box 1/1	Catalog Number	20006259
Part Number	20003464	Lot Number	A122798

### Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).

### Test Results

Metric	Specification	Result
ACD1 Concentration (ng/μl)	≥ 10	Pass
Slope of ACD1 standard Curve	$-3.8 \leq X \leq -3.2$	Pass
R2 of ACD1 Standard Curve	≥ 0.985	Pass
NTC	≥ 35.0	Pass

### Certification

This document certifies that the product(s) described above meet quality specifications.

### Quality Review

Print Name	Donna Mae Cruz	Signature		Date	16 AUG 2018
------------	----------------	-----------	--	------	-------------