

## Certificate of Analysis

## Description

Product Name	TruSeq Custom Amplicon Dx FFPE QC Box 1/1	Catalog Number	20006259
Part Number	20003464	Lot Number	A160498

## Test Conditions

The kitted reagents were tested for functional performance by qPCR amplification using control DNA (ACD1) and the included QC Primers (QCP).

## Test Results

Metric	Specification	Result
ACD1 Concentration (ng/ $\mu$ l)	$\geq 10$	Pass
Slope of ACD1 standard Curve	$-3.8 \leq X \leq -3.2$	Pass
R2 of ACD1 Standard Curve	$\geq 0.985$	Pass
NTC	$\geq 35.0$	Pass

## Certification

This document certifies that the product(s) described above meet quality specifications.

## Quality Review

Print Name	NAMALI PERERA	Signature		Date	18-AUG-2020
------------	---------------	-----------	--	------	-------------