
















Symbol Key

FOR IN VITRO DIAGNOSTIC USE

This document is a general legend to assist comprehension of symbols that can appear in the labeling for this product.

	For <i>in vitro</i> diagnostic use
	Catalog number - Indicates the manufacturer's catalog number for device identification
	Serial number
	Batch code - Indicates the manufacturer's batch code for batch or lot identification
	Authorized representative in the European Community
	Prescription only - US Federal Law restricts this device to sale by or on the order of a licensed practitioner
	High output
	European Conformity
	Consult instructions for use
	Temperature limit - Indicates acceptable shipping and exposure temperature limits
	Manufacturer - Indicates the medical device manufacturer, as defined in EU directives
	Use-by date
	Store away from sunlight - Indicates a medical device that requires protection from light or heat
	Health Hazard Symbol - For an explanation of the hazard symbols and the accompanying hazard statement, please refer to component Safety Data Sheet (SDS) which can be downloaded from http://support.illumina.com/sds.html
	Hazard Warning Symbol - For an explanation of the hazard symbols and the accompanying hazard statement, please refer to component Safety Data Sheet (SDS) which can be downloaded from http://support.illumina.com/sds.html

Copyright and Trademarks

© 2021 Illumina, Inc. All rights reserved. All trademarks are the property of Illumina, Inc. or their respective owners. For specific trademark information, see www.illumina.com/company/legal.html.

Contact Information



Illumina
 5200 Illumina Way
 San Diego, California 92122 U.S.A.
 +1.800.809.ILMN (4566)
 +1.858.202.4566 (outside North America)
techsupport@illumina.com
www.illumina.com



Illumina Netherlands B.V.
 Steenoven 19
 5626 DK Eindhoven
 The Netherlands

Australian Sponsor

Illumina Australia Pty Ltd
 Nursing Association Building
 Level 3, 535 Elizabeth Street
 Melbourne, VIC 3000 Australia