



Ref. Certif. No.

DE 2-035713

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT  
(IECEE) CB SCHEME

## CB TEST CERTIFICATE

|   |   |
|---|---|
| Product   | Genetic Sequencer   |
| Name and address of the applicant   | Illumina, Inc.<br>5200 Illumina Way<br>San Diego CA 92122, USA    |
| Name and address of the manufacturer                                      | Illumina, Inc.<br>5200 Illumina Way<br>San Diego CA 92122, USA    |
| Name and address of the factory   | Illumina, Inc.<br>25861 Industrial Blvd.<br>Hayward CA 94545, USA |
| Ratings and principal characteristics                                     | AC 100-240V, 50/60Hz; 400W; Class I                               |
| Trademark (if any)  | illumina  |
| Customer's Testing Facility (CTF) Stage used                              |   |
| Model / Type Ref.   | MiSeqDx, MiSeq, MiSeqFGx  |
| Additional information (if necessary may also be reported on page 2)      | Replaces CB Certificate DE 2031864.                               |
| A sample of the product was tested and found to be in conformity with     | IEC 61326-1:2020<br>for national differences see test report      |
| As shown in the Test Report Ref. No. which forms part of this Certificate | US21G3UW 004  |

This CB Test Certificate is issued by the National Certification Body



**TÜVRheinland**®

TÜV Rheinland LGA Products GmbH  
Tillystr. 2, 90431 Nürnberg, Germany  
Phone + 49 221 806-1371  
Fax + 49 221 806-3935  
Mail: cert-validity@de.tuv.com  
Web : www.tuv.com

Date: 2022-05-11

Signature:

Dipl.-Ing. M. Glagla