

Date : 2020/02/28

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Attn: Carol Escano

Re. : CU US + Canada Certificate

Type of Equipment : Next Generation Sequencer
Model Designation : See Certificate
Certificate No. : CU 72200420 0001
File No. : 32080459 002
Engineer/Contact : Ernesto Fernandez
Standard(s) : UL 61010-1:2012 R4.16
UL 61010-2-010:2015
UL 61010-2-081:2015
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2
CAN/CSA-C22.2 NO. 61010-2-010:15
CAN/CSA-C22.2 NO. 61010-2-081:15

Dear Ms. Escano,

The above referenced technical equipment has been tested and was found to be in compliance with the listed test requirement(s). Enclosed, please find the TUV Rheinland approval document No. CU 72200420 0001.

It authorizes you to label the listed product(s) with the TUV Rheinland Mark identified in the approval document. For compliance, the Test Mark must be on the approved

TUV Rheinland
of North America, Inc.
San Francisco Office
1279 Quarry Lane
Suite A
Pleasanton, CA 94566

Your product is subject to regular factory follow-up inspections as well as annual certificate and factory registration fees.

Tel 925-249-9123
Fax 925-249-9124
Mail info@us.tuv.com

In using the TUV Rheinland Mark you are obligated to comply with the TUV Rheinland of North America Service Agreement.

North American Headquarters

If we can be of any further assistance to you, please do not hesitate to contact us.

12 Commerce Road
Newtown, CT 06470

Sincerely yours,
Certification Body

Tel 203-426-0888
Fax 203-426-4009
Toll Free TUV-RHEINLAND
Web www.tuv.com

Dipl.-Ing. M. Glagla
QA Certification Officer

Member of
TUV Rheinland Group

Enclosure

Certificate



Certificate no.

CU 72200420 01

License Holder:

Illumina, Inc.
5200 Illumina Way
San Diego CA 92122
USA

Manufacturing Plant:

Illumina Singapore Pte. Ltd.
North Tech Lobby 3 #02-13/18
29 Woodlands Industrial Park E1
Singapore 757716
Singapore

Test report no.: USA-EF 32080459 002

Client Reference: Carol Escano

Tested to:

UL 61010-1:2012 R4.16
UL 61010-2-010:2015
UL 61010-2-081:2015
CAN/CSA-C22.2 NO. 61010-1-12 + GI1 + GI2
CAN/CSA-C22.2 NO. 61010-2-010:15
CAN/CSA-C22.2 NO. 61010-2-081:15

Certified Product: Next Generation Sequencer

License Fee - Units

Model Designation: NextSeq 1000, NextSeq 2000

Rated Voltage: AC 100-240V, 50/60Hz

Rated Power: 750W

Protection Class: I

Appendix: 1, 1-5

Licensed Test mark:



Date of Issue

(day/mo/yr)

28/02/2020