



Ref. Certif. No.

US-TUVR-010445-A2

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Genetic Sequencer

Name and address of the applicant

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122, USA

Name and address of the manufacturer

Illumina, Inc.  
5200 Illumina Way  
San Diego CA 92122, USA

Name and address of the factory

Illumina Singapore Pte. Ltd. North Tech Lobby 3 #02-13/18  
29 Woodlands Industrial Park E1  
Singapore 757716, Singapore

Ratings and principal characteristics

AC 100-240V, 50/60Hz; 600W  
Class I  
Laser Class 1

Trademark (if any)

illumina

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

NextSeq 500, NextSeq 550

Additional information (if necessary may also be reported on page 2)

Re-issuance of CB Certificate US-TUVR-10445 due to 2nd modification (update of critical component list). See associated test report for details.

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2010  
IEC 61010-2-010:2014  
IEC 61010-2-081:2015  
IEC 60825-1:2014  
for national deviations see test report

As shown in the Test Report Ref. No. which forms part of this Certificate

31393108 028

This CB Test Certificate is issued by the National Certification Body



TUV Rheinland of North America, Inc.  
12 Commerce Road  
Newtown, CT 06470 - USA

Date: 22.02.2018

Signature: Martin Glagla